

PRESSURE INFUSION BAG

MESH PANEL | CLEAR PANEL



Clear panel bags enables full visibility of solution bag to monitor fluid level and easily scan patient barcodes



FIGURE 3: 360° view of the color-coded gauge facilitates accurate pressure monitoring while the safety pressure relief valve prevents over-inflation

FIGURE 1: Solution bag hook enables easy bag changing without having to remove the infusion bag from the IV pole; keeps fluid bag in proper position while in use

FIGURE 2: Three-way stopcock ensures precise control of pressure while the large oval bulb allows for quick bladder inflation, all with ONE hand

8807, 8808, 8809 MESH PANEL
8810, 8811 CLEAR PANEL

INSTRUCTIONS FOR USE (IFU)

1. Remove pressure infusion bag from package, unfold pressure bag and hang pressure bag on to IV pole by the loop located at the top of the pressure bag (Figure 1).
2. Insert intravenous bag between mesh panel and pressure bag on 8807, 8808 and 8809 only.
3. If using 8810 or 8811, place intravenous bag between clear panel and pressure bag.
4. Hang loop of intravenous bag on to welded solution bag hook located just above panel on pressure bag (Figure 1).
5. FOR 8809 ONLY: Slide hanger loop through top of 3000 ml intravenous bag and rehang on IV pole.
6. Inflate pressure infusion bag by rotating the off portion of the valve/stopcock to the right. Pump inflation bulb until gauge reaches the proper pressure (300 mmHg) indicated in the green portion on the manometer (Figure 2).
7. Once proper pressure is reached on gauge, rotate off portion of the valve/stopcock straight up towards the pressure bag to maintain pressure in bag. **NOTE: Precise positioning of valve/stopcock is important or bag could lose pressure.**
8. To deflate bag turn the off portion on the valve/stopcock straight down towards inflation bulb to release pressure in bag.
9. Proper valve positioning is illustrated on all pressure bags.

All PremierPro™ Pressure Infusion Bags are recommended for single patient use.

CAUTION: All PremierPro™ Pressure Infusion bags come equipped with a pressure relief valve that prevents over inflation. The manometer will indicate red, along with the word “caution” when the valve reaches 330 mmHg. When the valve reaches 375 mmHg, pressure will release from the valve until it decreases below 375 mmHg (Figure 3)

WARNING: Continuous monitoring of the gauge pressure is required, as pressure may adjust from fluid being depleted from the fluid bag. inflate as needed, ensuring not to over inflate.



MRI Safety Information: The PremierPro™ Pressure Infusion Bag is MR Conditional under the following conditions.

- Name of the Device: PremierPro™ Pressure Infusion Bag
- Static Magnetic Field (T): 7-T or less
- Maximum Spatial Field Gradient (T/m and gauss/cm): 19-T/m (1,900-gauss/cm)